

Return & Repair Form

Customer Information:

CUSTOMER NJ STATE POLICE Cust # _____
 Date of Return: 10/12 Carrier: Fed Ex Method: 2ND

Product Information:

Product: 7110 7410 6510 6810 Serial # FRNK 0050
 Description: A B PLUS SCREENER _ Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # _____

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter Regulator Mag Card Rdr # _____
- Printer Paper Printer Ribbon Casio # _____
- Mouthpieces Carrying Case Dry Gas
- Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: Error 09

Part #	Description	Qty	Total
<u>WMPLABOR</u>	<u>LABOR</u>	<u>.5</u>	<u>N/C</u>
<u>WMPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	<u>N/C</u>

Repair Notes: REPLACED KL RELAY SWITCH ON MOTHER BOARD
W/ NON INVENTORY ITEM. CAL OPS CHECK + QC.

Technician: Art Dly

Date: 10/25/05