RETURN & REPAIR FORM

CUSTOMER N.J STATE POLICE Cust. # 1005
Acc Exp 2004

Date ret. 10/5 Carrier UPS Method 2nd RMA# _____

7110 Serial# ARNL 0057 Warranty exp.: _____

Reported Problem: Jumpy IR

Accessories (check all that apply): ☐ Regulator ☐ P.Paper ☐ Scotty V ☐ PrinterRib. ☐ Pouch
Sim.Temp.Probe Ser.# _____ ☐ Keyboard ☐ Organizer ☐ Key
☐ Mouthpieces_____ ☐ Sim. to cuvette hose ☐ Pump to Sim. hose ☐ Power cord
☐ Other (specify) _____
☐ Other (specify) _____

Part#	Description	$	Qty	Total
WD 8311806	IR-DETECTOR		1	
* CAL71			1	
		Repair Time		Time $

TOTAL REPAIR $ _____

REPAIR NOTES:
Replaced IR-DETECTOR. CALIBRATED UNIT full QC & OPS CHK.

Date 09092002 Technician [illegible]
Every above checked item is included ☐ Shipping _____

G:\HOME\COMMON\FORMS\RRFORM.WPD