

RETURN & REPAIR FORM



CUSTOMER NE State Police Cust. # 72477612
 Date ret. 5/25 Carrier UPS Method G11 RMA#
 7110 Serial# ARNL-0051 Warranty exp.:

Reported Problem:

Accessories (check all that apply): ☐ Regulator ☐ P.Paper ☐ Scotty V ☐ Printer Rib. ☐ Pouch
 Sim. Temp. Probe Ser.# ☐ Keyboard ☐ Organizer ☐ Key
☐ Mouthpieces ☐ Sim. To cuvette hose ☐ Pump to Sim. hose ☐ Power cord
☐ Other (specify)
☐ Other (specify)

<u>Part #</u>	<u>Description</u>	<u>\$</u>	<u>Qty</u>	<u>Total</u>
6808455	Fuel Cell		1	
MPAL 71			1	
		Repair Time	.5	Time \$

TOTAL REPAIR \$

REPAIR NOTES:
Replaced fuel cell. Upgraded Unit to NJ-3.10.
Full QC & OAS CHK WITH CALIBRATION.

Date 19 OCT 04 Technician EW
 Any items needed from customer or shipping