

Return & Repair Form:

Customer Information: **KEVIN F.**

Company Name: MT ADTU B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 3.3.09 Date Given to Service: 3.4.09  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY Other

Product Information:

Product: 7410 - 7110 - 6510 - 6810 Serial #: AR NL - 0051  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR  
 Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P - \_\_\_\_\_

ACCESSORIES:

110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

Repair Information:

Test #


Reason for Return: ERROR 112

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>*</u>
<u>MPCAR 71</u>	<u>CYL</u>	<u>1</u>	
<u>MPCARBOR</u>	<u>LABOR</u>	<u>.5</u>	

Repair Notes:

CLEAR ERROR 112 / FC IS OF (6-2004) REPLACED  
CAL W/ QC AND OPS CHECKS

Service Technician



Date: 03/04/2009