

# Return & Repair Form:

**Customer Information:** KEVIN F.

Company Name: NT ADTU B: \_\_\_\_\_ S: \_\_\_\_\_

Date Received: 3.3.09 Date Given to Service: 3.4.09

Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY Other

**Product Information:**

Product: 7410 - 7110 - 6510 - 6810 Serial #: AR NL - 0051

Description: A - B - Pl \_\_\_\_\_

Whole Instrument \_\_\_\_\_

Other \_\_\_\_\_

**ACCESSORIES:**

110 V A/C Adapter

Printer Paper

Mouthpieces

Other (Please Specify) \_\_\_\_\_

BILLABLE  
REPAIR

**Repair Information:**

Test # \_\_\_\_\_

Reason for Return: ERROR 112

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>*</u>
<u>MPCTR 71</u>	<u>CTR</u>	<u>1</u>	
<u>MPCTR LABOR</u>	<u>LABOR</u>	<u>.5</u>	

Repair Notes:

CLEAR(ED) ERROR 112 / FC IS WF (6-2004) REPLACED  
CTR W/ QC AND OPS CHECKS

Service Technician \_\_\_\_\_

Date: 03/04/2009