

# Return & Repair Form:

## Customer Information:

Company Name: Don Paulist / NJSP B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 1-19-10 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR NL -0051  
 9510- Drug Tester 5000  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top ½ Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return:

Part Number	Description	Qty	Total Cost
<u>MISC PART</u>	<u>SWITCH 7110</u>	<u>1</u>	<u>N/C WORK</u>
<u>M/Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C WORK</u>

## Repair Notes:

Unit came in with a problem with power not up, unit would not turn on, there was not any power at the 110 in-let, replace in-let unit power up.

I Ran a STD - check # 2 and Diagnostic check

Service Technician: AS

Date: 1/22/2010