


RETURN & REPAIR FORM



CUSTOMER NJ State Police Cust. # _____

Date ret. 5/26 Carrier UPS Method G1D RMA# _____

7110 Serial# ARNL-0052 Warranty exp.: _____

Reported Problem: _____

- Accessories (check all that apply): Regulator P.Paper Scotty V Printer Rib. Pouch
 Sim. Temp. Probe Ser.# _____ Keyboard Organizer Key
 Mouthpieces _____ Sim. To cuvette hose Pump to Sim. hose Power cord
 Other (specify) _____
 Other (specify) _____

<u>Part #</u>	<u>Description</u>	<u>\$</u>	<u>Qty</u>	<u>Total</u>
<u>6708455</u>	<u>Fuel Cell</u>		<u>1</u>	
<u>MPCAL 71</u>			<u>1</u>	
		Repair Time	<u>.5</u>	Time \$

TOTAL REPAIR \$ _____

REPAIR NOTES:
Replaced fuel cell, upgraded unit to NS-3.10
Full QC & ORS CHK WITH CALIBRATION.

Date 19 OCT 04 Technician [Signature]

Every above checked item is included Shipping _____