

Return & Repair Form:

Customer Information:

Company Name: NOSP/DOO B: _____ S: _____
 Date Received: 11-19-08 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR NL - 0054
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: FC

| Part Number | Description | Qty | Total Cost |
|-----------------|--------------|------------|-----------------|
| <u>6808455</u> | <u>FC</u> | <u>1</u> | <u>N/C W/WH</u> |
| <u>MPCAL 71</u> | <u>71</u> | <u>1</u> | <u>N/C W/WH</u> |
| | | | |
| | | | |
| | | | |
| <u>MP Labor</u> | <u>Labor</u> | <u>1.5</u> | <u>N/C W/WH</u> |

Repair Notes:

FC MR UF-0763 replaced with FC MR HJ-0258

Cal w/DC ; ops check

Service Technician: AS

Date: 12/01/09