


RETURN & REPAIR FORM



ALCOTEST 7110


CUSTOMER NJ State Police Cust. # _____

Date ret. 5/25 Carrier UPS Method G1D RMA# _____

7110 Serial# ARNC 0056 Warranty exp.: _____

Reported Problem: _____

- Accessories (check all that apply):
- Regulator
 - P.Paper
 - Scotty V
 - Printer Rib.
 - Pouch
 - Sim. Temp. Probe Ser.# _____
 - Keyboard
 - Organizer
 - Key
 - Mouthpieces _____
 - Sim. To cuvette hose
 - Pump to Sim. hose
 - Power cord
 - Other (specify) _____
 - Other (specify) _____

<u>Part#</u>	<u>Description</u>	<u>\$</u>	<u>Qty</u>	<u>Total</u>
	<u>Fuel Cell</u>		<u>1</u>	
<u>MPCAL 71</u>			<u>1</u>	
		Repair Time	<u>.5</u>	Time \$

TOTAL REPAIR \$ _____

REPAIR NOTES:
Replaced Fuel Cell. Calibrated Unit. full ac
& OPS CHK.

Date 20 Jul 04 Technician [Signature]

Every above checked item is included Shipping _____