

Return & Repair Form

Customer Information:

CUSTOMER NJ STATE POLICE Cust # _____

Date of Return: 10/12 Carrier: Fed Ex Method: 2ND

Product Information:

Product: 110 7410 6510 6810 Serial # ARNL-0056

Description: A B PLUS SCREENER Printer Serial # _____

Whole Inst. Top Half Simulator Serial # _____

OTHER: _____ Probe Serial # _____

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter
- Regulator
- Mag Card Rdr # _____
- Printer Paper
- Printer Ribbon
- Casio # _____
- Mouthpieces
- Carrying Case
- Dry Gas
- Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: ERROR 75

Part #	Description	Qty	Total
<u>W MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	<u>N/C</u>
<u>W MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	<u>N/C</u>

Repair Notes: RESOLDERED LEAD ON BREATH HOSE.
Cal OPS CHECK + QC.

Technician: Chris Dwy

Date: 10/24/05