

RETURN & REPAIR FORM



CUSTOMER NJ State Police Cust. # 103/44/610  
 Date ret. 5/25 Carrier UPS Method G-11 RMA# \_\_\_\_\_  
 7110 Serial# ARNL-0064 Warranty exp.: \_\_\_\_\_

Reported Problem: \_\_\_\_\_

- Accessories (check all that apply):  Regulator  P.Paper  Scotty V  Printer Rib.  Pouch  
 Sim. Temp. Probe Ser.# \_\_\_\_\_  Keyboard  Organizer  Key  
 Mouthpieces \_\_\_\_\_  Sim. To cuvette hose  Pump to Sim. hose  Power cord  
 Other (specify) \_\_\_\_\_  
 Other (specify) \_\_\_\_\_

<u>Part #</u>	<u>Description</u>	<u>S</u>	<u>Qty</u>	<u>Total</u>
6608455	Fuel Cell		1	
MPCAL 71			1	
		Repair Time	.5	Time \$

TOTAL REPAIR \$ \_\_\_\_\_

REPAIR NOTES:  
 Replaced fuel cell, Upgraded Unit to NJ-3.10.  
 full Q&A OPS CHK WITH CALIBRATION.

Date 19 OCT 04 Technician [Signature]  
 Company name (check item returned) \_\_\_\_\_