

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER Woodbridge Police Cust # _____
 Date of Return: 9/26/10 Carrier: Fed Ex Method: standard

Product Information:

Product: 7110 7410 6510 6810 Serial # ARRL-0019
 Description: A B PLUS SCREENER _ Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____
 Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: fuel cell replacement

Part #	Description	Qty	Total
6808455	Fuel Cell	1	
MPLABOR	LABOR	.5	
MPLACT1	7110 CAL	1	

Repair Notes: Replaced leaky fuel cell. Full cal ops check + Q.E.

Technician: CRD

Date: 9/27/10