

Return & Repair Form:

Customer Information:

Company Name: Wood Bridge Police B: _____ S: _____
 Date Received: 1-7-10 Date Given to Service: _____
 Carrier: (FedEx) - UPS - DHL - USPS Method: GRD - (NDA) - 2DY - 3DY - Other

Product Information:

Product: 7410-(7110)-6510-6810-7510 Serial #: AR ^{Priority} RL - 0019
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top ½ Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: FC

Part Number	Description	Qty	Total Cost
<u>600 8453</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP CGL 71</u>	<u>CGL</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C WARR</u>

Repair Notes:

NEW FC # ARAJ-0136
CGL w/ QC & Ops Check

Service Technician: JS

Date: 01/11/2010