

# RETURN AND REPAIR FORM



## Customer Information

B: \_\_\_\_\_ S: \_\_\_\_\_

Company Name: Woodbridge P.D., Woodbridge, NJ

Date Received: 08/16/2016

Date given to service: 08/16/2016

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR RL-0019

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: December 2016

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

## Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	4	NC/W
6808455	Fuel Cell	1	NC/W
6808374	Geared Engine	1	NC/W

Repair Notes: Replaced defective fuel cell and geared engine.

Was unable to duplicate customer complaint of orange button not working after testing it for 2 weeks and unit stuck in searching mode.

Fuel Cell = ARJC-1498

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 09/20/2016