

# Dräger safety

## Return & Repair Form

### Customer Information:

CUSTOMER O'Lawrence River Port AUTH Cust # \_\_\_\_\_  
 Date of Return: 3/22 Carrier: Fed Exp Method: PRIORITY

### Product Information:

Product: 7110 7410 6510 6810 Serial # APML-0048  
 Description: A B PLUS SCREENER \_ Printer Serial # \_\_\_\_\_  
 Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
 OTHER: \_\_\_\_\_ Probe Serial # \_\_\_\_\_  
 Demo Unit  Returned to stock on: \_\_\_\_\_  
 Accessories: (check all that apply):  
 110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (specify): \_\_\_\_\_  
 Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: Error 075

Part #	Description	Qty	Total
MPLB012	LABOR	.5	
MPCAL11	7110 CAL	1	

Repair Notes: Soldered connection on breath hose.  
Cal 225 check + Q.C.

Technician: [Signature]

Date: 3/23/06