

Return & Repair Form

Customer Information:

CUSTOMER MIDDLE TWP P.D. NJ Cust # _____
 Date of Return: 12/12/2007 Carrier: (FX) UPS DHL USPS Method: GRD NDA 2DY 3DY
 OTHER: STUD O/N

Product Information:

Product : (7110) 7410 6510 6810 Serial # AR RN - 0077
 Description: A B PLUS SCREENER _ Printer Serial # AR -
 Whole Inst. Top Half Simulator Serial # DDXC53-0139
 OTHER : _____ Probe Serial # DDWSPZ-300
 Demo Unit Returned to stock on: _____
 Accessories: (check all that apply):
 Regulator Mag Card Rdr # _____
 Printer Ribbon Casio # _____
 Carrying Case Dry Gas
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: _____

Part #	Description	Qty	Total
<u>MPCERT</u>	<u>SIM CERT</u>	<u>1</u>	<u>WARR N/C</u>
<u>MPCALTY</u>	<u>PROBE CAL</u>	<u>1</u>	<u>WARR N/C</u>
<u>MPCAL 71</u>	<u>7110 CAL</u>	<u>1</u>	<u>WARR N/C</u>
<u>680B455</u>	<u>FC</u>	<u>1</u>	<u>WARR N/C</u>
<u>MP LABOR</u>	<u>LABOR</u>	<u>1.0</u>	<u>WARR N/C</u>

Repair Notes: "PROBE" WAS SET @ 101 IN 7110 - USED CUSTOMER
PROBE (106) & SIMULATOR TO TEST TEMP = READ 34.19
DO AUTO ADJUST ON CIRCUIT - NOW READS 34.03 AT OFFSET @ 1847/218
FC HAD CORROSION - REPLACED CAL W/ QC & OPS CHECKS

Technician: [Signature]

Date: 12.13.07