

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER Stafford Twp Police Cust # 1500-51459
 Date of Return: 02/15/2008 Carrier: UPS DHL USPS Method: GRD NDA 2DY 3DY
 OTHER: _____

Product Information:

Product : 7110 6510 6810 Serial # ARSA -0031
 Description: A B PLUS SCREENER Printer Serial # AR -
 Whole Inst Top Half Simulator Serial # _____
 OTHER : _____ Probe Serial # DD P -

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter Regulator Mag Card Rdr # _____
- Printer Paper Printer Ribbon Casio # _____
- Mouthpieces Carrying Case Dry Gas
- Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: _____

| Part # | Description | Qty | Total |
|---------|-------------|-----|---------|
| 6808486 | PLATE | 1 | w/cwarr |
| 6808455 | FC | 1 | w/cwarr |
| MPCAC H | 7110 CAC | 1 | w/cwarr |
| MPLABOR | LABOR | .5 | w/cwarr |

Repair Notes: REPLACE FC / CAC w/ QC & OPS CHECKS
REPLACED 1 FC PLATE

Technician: [Signature]

Date: 02-21-08