

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Metuchen Boro P.D., Metuchen, NJ

Date Received: 02/04/15 Date given to service: 02/04/15

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810 Serial Number: AR SA-0137
 7110 8610 Printer Serial#: AR
 7510 DT5000 Sim Serial#: DD
 7410 Upper-half 9510 Probe Serial#: DD
 7410 Whole

Warranty Expires: January 2016

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	1	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W
6808374	Geared Engine Motor	1	NC/W

Repair Notes: Replaced defective geared engine motor, fuel cell, and spacer plates.

Unit was giving Error 32, geared engine motor was changed.

Fuel Cell = ARFJ-0918

CAL W/QC AND OPS CHECK

Service Technician BC Date: 02/13/15