

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Metuchen P.D., Metuchen, NJ

Date Received: 11/10/2017

Date given to service: 11/10/2017

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR SA-0137

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

| Part Number | Description | Qty | Total Cost |
|-------------|-------------|-----|------------|
| 4414161 | Calibration | 1 | |
| 4414166 | Labor | 1 | |
| 6808455 | Fuel Cell | 1 | |
| | | | |
| | | | |
| | | | |
| | | | |

Repair Notes: Replaced defective fuel cell.

Adjusted Service CH 3, was too low and resolved the BLOWING NOT ALLOWED issue.

FUEL CELL = ARKA-1711

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 11/14/2017