



Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER Glassboro Police Dept. Cust # 50682
 Date of Return: 9/15/06 Carrier: Fed Ex Method: Priority

Product Information:

Product: 7110 7410 6510 6810 Serial # ARSB-0082
 Description: A B PLUS SCREENER _ Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____
 Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

** Not packaged very well*

Repair Information:

Test # _____

Reason for Return: Low EC Readings

Part #	Description	Qty	Total
<u>6808455</u>	<u>Fuel Cell</u>	<u>1</u>	
<u>MPLABOR</u>	<u>LABOR</u>	<u>1.5</u>	
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Confirmed low EC readings. Inspection of F/C revealed it had leaked. Replaced F/C. Full cal ops check + O.C.

Technician: CRD

Date: 9/18/06