

# Dräger safety

## Return & Repair Form

### Customer Information:

CUSTOMER HopATcong P.O Cust # 50192  
 Date of Return: 4/17 Carrier: Fed Ex Method: STB

### Product Information:

Product : 7110 7410 6510 6810 Serial # FRSB-0092  
 Description: A B PLUS SCREENER \_ Printer Serial # \_\_\_\_\_  
 Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
 OTHER : \_\_\_\_\_ Probe Serial # \_\_\_\_\_  
 Demo Unit  Returned to stock on: \_\_\_\_\_  
 Accessories: (check all that apply):  
 110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (specify): \_\_\_\_\_  
 Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: Couldnt upgrade firmware (Invalid Packet size)

Part #	Description	Qty	Total
MPCAL71	7110 CAL	1	N/C
B315075	Mother Board	1	N/C
MPMISC.	Cuvette End Covers	2	N/C
	<b>AXED</b>		
	<u>4/19/06 2:50PM</u>		

Repair Notes: Observed damage to unit. Modem was loose, Display was misaligned, Case bent in several places, P/S cover was ajar, both end covers on cuvette broken, internal chassis bent, security slug bent. Replaced cuvette ends and straightened bent parts. Replaced M/B to fix comm problem.

Technician: RD

Date: 4/19/06