

# RETURN AND REPAIR FORM



**Customer Information**

B: \_\_\_\_\_ S: \_\_\_\_\_

Company Name: Hopatcong P.D., Hopatcong NJ

Date Received: 12/16/11 Date given to service: \_\_\_\_\_

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610 Serial Number: AR SB-0092  
 7510  DT5000 Printer Serial#: AR  
 7410 Upper-half  9510 Sim Serial#: DD  
 7410 Whole Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

**Accessories**

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

**Repair Information:**

Test#: \_\_\_\_\_

Part Number	Description	Qty	Total Cost
<u>12029</u>	<u>Power Supply</u>	<u>1</u>	
<u>mp cal 71</u>	<u>Calibration</u>	<u>1</u>	
<u>mp labor</u>	<u>Labor</u>	<u>1.5</u>	

Repair Notes: Install a new power supply, because  
unit would not power up.

Cal w/DC; ops check

Service Technician: [Signature]

Date: 12-19-12