

# Dräger safety

## Return & Repair Form

### Customer Information:

CUSTOMER MANCHESTER TWP POLICE Cust # \_\_\_\_\_

Date of Return: 21 B/2008 Carrier  UPS  DHL  USPS Method:  GRD  NDA  2DY  3DY  
OTHER: \_\_\_\_\_

### Product Information:

Product:  7110 7410 6510 6810 Serial # AR SB - 0095

Description: A B PLUS SCREENER Printer Serial # AR -

Whole Inst. Top Half Simulator Serial # \_\_\_\_\_

OTHER: \_\_\_\_\_ Probe Serial # DD P -

Demo Unit  Returned to stock on: \_\_\_\_\_

### Accessories: (check all that apply):

110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_

Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_

Mouthpieces  Carrying Case  Dry Gas

Other (specify): \_\_\_\_\_

Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: FC

| Part #          | Description     | Qty       | Total           |
|-----------------|-----------------|-----------|-----------------|
| <u>6808455</u>  | <u>FC</u>       | <u>1</u>  | <u>N/C WARR</u> |
| <u>MPCAL 71</u> | <u>7110 CAL</u> | <u>1</u>  | <u>N/C WARR</u> |
|                 |                 |           |                 |
|                 |                 |           |                 |
| <u>MPLABOR</u>  | <u>LABOR</u>    | <u>.5</u> | <u>N/C WARR</u> |

Repair Notes: REPLACE FC / CAL w/ QC & OPS CHECKS

Technician: [Signature]

Date: 02-09-08