

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Manchester TWP P.D., Manchester, NJ

Date Received: 02/15/13

Date given to service: 02/15/13

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR SB-0095

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: January 2014

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plate	1	NC/W
MP Labor	Labor	1	NC/W

Repair Notes: Replaced defective fuel cell and spacer plate.

Fuel Cell ARDM-2230

Replaced defective C156 and C167 = modem line busy

CAL W/QC AND OPS CHECK

Service Technician: BC

Date: 02/21/13