

Return & Repair Form:

Customer Information:

Company Name: W/Brookfield PD B: 81012 S: 48620
 Date Received: 04/20/2009 Date Given to Service: 04/20/2010
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 7110 - 6510 - 6810 Serial #: AR SB - 0110
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -

ACCESSORIES:

110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: FC / PLATE

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	
<u>6808486</u>	<u>Plate</u>	<u>1</u>	
<u>MP CCL 41</u>	<u>CCL</u>	<u>1</u>	
<u>MP Labor</u>	<u>Labor</u>	<u>1hr</u>	

Repair Notes:

Replace FC and Plate
CAL w/AC & OPS Check

Service Technician: [Signature]

Date: 04/20/2009