

# Return & Repair Form:

## Customer Information:

Company Name: Comm. Barry Bridge Police B: \_\_\_\_\_ S: \_\_\_\_\_

Date Received: 1-4-11 Date Given to Service: \_\_\_\_\_

Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410-7110-6510-6810-7510  
9510- Drug Tester 5000

Serial #: AR SC - 0023

Description: A - B - Plus - Screener - Demo

Printer Ser #: AR -

Whole Instrument

Top 1/2

Sim Ser #: \_\_\_\_\_

Other \_\_\_\_\_

Probe Ser # DD P -

## ACCESSORIES:

110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_

Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_

Mouthpieces  Carrying Case  Dry Gas

Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

## Test #

Reason for Return: FC

Part Number	Description	Pr	Qty	Total Cost
<u>10808455</u>	<u>FC</u>		<u>1</u>	<u>N/C W</u>
<u>MP CAL 71</u>	<u>CAL</u>		<u>1</u>	<u>N/C W</u>
<u>MP Labor</u>	<u>Labor</u>		<u>1.5</u>	<u>N/C W</u>

Repair Notes:

FC AR513-0210 Replaced with FC #  
AR135-1824

CAL W/ QC & OPS check

Service Technician: [Signature]

Date: 01/05/2011