

Return & Repair Form:

1044764 / 80583904

Customer Information:

Company Name: MT. Laurel P.D. B: 47613

Date Received: 3/3/2011 Date Given to Service: 3/3/2011

Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7440-7110-6510-6810-7510 Serial #: AR 5C - 0047
 9510- Drug Tester 5000

Description: A - B - Plus - Screener - Demo Printer Ser #: AR

Whole Instrument Top 1/2 Sim Ser #: _____

Other _____ Probe Ser # DD P

ACCESSORIES:

110 V A/C Adapter Regulator Mag Card Rdr# _____

Printer Paper Printer Ribbon Casio # _____

Mouthpieces Carrying Case Dry Gas

Other (Please Specify) _____ Warranty Exp. Date Oct 2013

Repair Information:

Test #

Reason for Return:			
<u>FC</u>			
Part Number	Description	Qty	Total Cost
<u>6808453</u>	<u>FC</u>	<u>1</u>	<u>n/c w</u>
<u>MP CAL 71</u>	<u>CAL</u>	<u>1</u>	<u>n/c w</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>n/c w</u>

Repair Notes:

FC # AR 13W-0873 replaced FC #
AR KN-0517

Cal w/DC 5 ops check

Service Technician: [Signature]

Date: 3/4/2011