

## Return & Repair Form

### Customer Information:

CUSTOMER DELAWARE RIVER PORT AUTH. POLICE Cust # \_\_\_\_\_  
 Date of Return: 3/12/2008 Carrier  UPS  DHL  USPS Method:  GRD  NDA  2DY  3DY  
 OTHER: \_\_\_\_\_

### Product Information:

Product : 7110 7410 6510 6810 Serial # AR5C-0052  
 Description: A B PLUS SCREENER Printer Serial # AR -  
 Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
 OTHER : \_\_\_\_\_ Probe Serial # DD P -  
 Demo Unit  Returned to stock on: \_\_\_\_\_  
 Accessories: (check all that apply):  
 110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (specify): \_\_\_\_\_  
 Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: FC + IR?

Part #	Description	Qty	Total
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>6808486</u>	<u>PLATE</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPCAL71</u>	<u>F110 CAL</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPLABOR</u>	<u>LABOR</u>	<u>-5</u>	<u>N/C WARR</u>

Repair Notes: REPLACE FC + PLATE / CAL W/QC & OPS CHECKS

Technician: [Signature]

Date: 03-12-08