

Return & Repair Form:

Customer Information:

Company Name: DRPA Wake Whitman ^{DR} B: 5/6/10 S: 5/10/8

Date Received: 9-21-10 Date Given to Service: _____

Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR SC-0052
 9510- Drug Tester 5000

Description: A - B - Plus - Screener - Demo Printer Ser #: AR

Whole Instrument Top ½ Sim Ser #: _____

Other _____ Probe Ser # DD P

ACCESSORIES:

110 V A/C Adapter Regulator Mag Card Rdr# _____

Printer Paper Printer Ribbon Casio # _____

Mouthpieces Carrying Case Dry Gas

Other (Please Specify) _____ Warranty Exp. Date April 2013

Repair Information:

Test #

Reason for Return: PCIB

Part Number	Description	Qty	Total Cost
<u>MP Cal 71</u>	<u>Cal</u>	<u>1</u>	<u>N/C w</u>
<u>8315075</u>	<u>PCIB</u>	<u>1</u>	<u>N/C w</u>
<u>MP Labor</u>	<u>Labor</u>	<u>1hr</u>	<u>N/C w</u>

Repair Notes:

Unit would not store data, replaced PCIB.

Cal w/ DC; ops check

Service Technician: [Signature]

Date: 9/27/2010