

Return & Repair Form:

Customer Information:

Company Name: DELAWARE RIVER PORT B: _____ S: _____
 Date Received: 10.12.10 ^{Auth.} Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR SC - 0052
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR
 Whole Instrument Top ½ Sim Ser #: _____
 Other _____ Probe Ser # DD P
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date April 2013

Repair Information:

Test #

Reason for Return:			
Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C W</u>
<u>MP Cal 71</u>	<u>Cal</u>	<u>1</u>	<u>N/C W</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C W</u>

Repair Notes:

JOE TESTED FUNCTION ACCESS WITH BLACK KEY, YELLOW KEY AND VIA MAZIO PROGRAM SERVICE CH 6 WITH BLACK KEY. ~100 ACCESS TEST WERE PERFORMED WITH NO FAILURE.

FC AR YE-0645 Dated May 2007 Replaced with FC AR BE-1014 installed for preventive maintenance.
Cal w/RC & Ops Check

Service Technician: [Signature]

Date: 10/15/2010