

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: D.E. River Port Auth, P.D., Camden, NJ

Date Received: 11/20/12

Date given to service: 11/20/12

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR SC-0052

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

Test#:

| Part Number | Description | Qty | Total Cost |
|-------------|-------------|-----|------------|
| MP Cal 71   | Calibration | 1   | NC/W       |
| 8315075     | PCB         | 1   | NC/C       |
| MP Labor    | Labor       | .5  | NC/W       |

Repair Notes: Instrument was missing a sequential file number, replaced (PCB).

CAL W/QC AND OPS CHECK

Service Technician JS

Date: 11/30/12