

RETURN AND REPAIR FORM



Customer Information

B: _____ S: _____

Company Name: DRPA P.D., Camden, NJ

Date Received: 03/19/2018

Date given to service: 03/19/2018

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR SC-0052

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#: _____

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	N/C
4414166	Labor	6	N/C
6808455	Fuel Cell	1	N/C
6808550	EC Sampling System	1	N/C

Repair Notes: Replaced defective fuel cell and ec sampling system.

Fuel Cell = ARKL-1111

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 03/20/2018