

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER BURLTON city Bridge Comm. Cust # 47658
 Date of Return: 4-13-7 Carrier: UPS Method: next Day Air

Product Information:

Product: 7110 7410 6510 6810 Serial # ARSC-0056
Description: A B PLUS SCREENER Printer Serial # _____
Whole Inst. Top Half Simulator Serial # _____
OTHER: _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____
Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: ? Random IR results

Part #	Description	Qty	Total
<u>6308455</u>	<u>Fuel Cell</u>	<u>1</u>	
<u>MPMISC</u>	<u>Preamp bd.</u>	<u>1</u>	<u>N/C</u>
<u>MPLABOR</u>	<u>LABOR</u>	<u>3</u>	
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Swapped preamp bds with another unit and and IR problem followed the preamp bd. Replaced preamp. Replaced leaky F/C. Full cal ops check + Q.C.

Technician: CRD

Date: 4/24/07