

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER Winslow Twp P.D. Cust # _____
 Date of Return: 3/13/07 Carrier: FedEx Method: Priority

Product Information:

Product: (7110) 7410 6510 6810 Serial # ARTL-0021
Description: A B PLUS SCREENER Printer Serial # _____
Whole Inst. Top Half Simulator Serial # _____
OTHER: _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____
Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: EC Problem

Part #	Description	Qty	Total
6808455	Fuel Cell	1	
MPLABOR	LABOR	.5	
MPLACT1	7110 CAL	1	

Repair Notes: Replaced leaking FIC. Full cal ops check and Q.C. #

Technician: CRD

Date: 3/13/07