

# Dräger safety

## Return & Repair Form

### Customer Information:

CUSTOMER Bridge water Twp. Police Cust # \_\_\_\_\_  
 Date of Return: 3/30/17 Carrier: FEDEX Method: Priority

### Product Information:

**Product:** 7110 7410 6510 6810 Serial # ARTL-0023  
**Description:** A B PLUS SCREENER Printer Serial # \_\_\_\_\_  
**Whole Inst. Top Half** Simulator Serial # \_\_\_\_\_  
**OTHER:** \_\_\_\_\_ Probe Serial # \_\_\_\_\_  
 Demo Unit  Returned to stock on: \_\_\_\_\_  
**Accessories:** (check all that apply):  
 110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (specify): \_\_\_\_\_  
 Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: EC Problem

| Part #         | Description      | Qty       | Total |
|----------------|------------------|-----------|-------|
| <u>6808455</u> | <u>Fuel Cell</u> | <u>1</u>  |       |
| <u>MPLABOR</u> | <u>LABOR</u>     | <u>.5</u> |       |
| <u>MPCAL71</u> | <u>7110 CAL</u>  | <u>1</u>  |       |
|                |                  |           |       |
|                |                  |           |       |

Repair Notes: Replaced leaky FLC. Full cal  
ops check + Q.C.

Technician: CRD

Date: 4/2/07