

Return & Repair Form:

Customer Information:

Company Name: BRIDGEWATER TWP P.D. B: _____ S: _____
 Date Received: 12-12-08 Date Given to Service: 12-12-08
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY 3DY - Other

Product Information:

Product: 7410 7110 - 6510 - 6810 Serial #: ARTL - 0023
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top ½ Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: LOW EC RESULTS.

Part Number	Description	Qty	Total Cost
<u>6B0B455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPCAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C WARR</u>
<u>mPLABOR</u>	<u>LABOR</u>	<u>.5</u>	<u>N/C WARR</u>

Repair Notes:

REPLACED FC AS PM - DATA DOES NOT SUPPORT REPLACEMENT -
INSPECT CUVETTE - DETECTOR - FILTER - ALL OK

CAL W/ QC & OPS CHECKS

Service Technician: [Signature]

Date: 12-15-2008