

Return & Repair Form:

Customer Information:

Company Name: Budgetmaster Temp B: _____ S: _____
 Date Received: 3-10-11 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-7110-6510-6810-7510 Standard Serial #: AR TL - 0023
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top ½ Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return:			
Part Number	Description	Qty	Total Cost
<u>mp Cal. 71</u>	<u>Cal</u>	<u>1</u>	<u>n/c w</u>
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>n/c w</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>n/c w</u>

Repair Notes:

Unit was test by programming 99 wet .1% test
the interval test printout did not show any
sign that there is a FC problem, spoke with
supervisor and he said he thinks it a environment
problem at the police station.

Cal w/ FC & ops check

ARCA-0178
 Service Technician: [Signature]

Date: 3/18/2011