

# Return & Repair Form:

## Customer Information:

Company Name: Hightstown B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 3/25/2010 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY Other \_\_\_\_\_

## Product Information:

Product: 7410-7110-6510-6810-7510  
 9510- Drug Tester 5000  
 Description: A - B - Plus - Screener - Demo  
 Serial #: AR TN - 0042  
 Printer Ser #: AR -  
 Whole Instrument \_\_\_\_\_ Top 1/2 \_\_\_\_\_  
 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD \_\_\_\_\_ P -  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return:

FC

Part Number	Description	Qty	Total Cost
<u>6808486</u>	<u>Plate</u>	<u>1</u>	<u>w/c warr</u>
<u>6808455</u>	<u>PL</u>	<u>1</u>	<u>w/c warr</u>
<u>mp Cal 71</u>	<u>Cal</u>	<u>1</u>	<u>w/c warr</u>
<u>mp Labor</u>	<u>Labor</u>	<u>1.5</u>	<u>w/c warr</u>

Repair Notes:

NEW FC# ARAN-0925

Cal w/DC & OPS check

Service Technician: JS

Date: 03/25/2010