

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER Lower Alloways Creek PD Cust # _____
 Date of Return: 12/16 Carrier: Fed Exp Method: Priority

Product Information:

Product : 7110 7410 6510 6810 Serial # ARTN-0048
 Description: A B PLUS SCREENER _____ Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER : _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____
 Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: Printer problem / alcohol showed error

Part #	Description	Qty	Total

Repair Notes: Resoldered leads on motor/sensor. Cal ops check + O.S.

Technician: CRD

Date: 12/19/05