

Return & Repair Form:

Customer Information:

Company Name: Carneys Point B: _____ S: _____
 Date Received: 2-3-09 Date Given to Service: _____
 Carrier: RedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 - 7110 - 6510 - 6810 Serial #: AR TN - 0055
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD _____ P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: ERROR 35

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C W/ARR</u>
<u>MPCAL 71</u>	<u>CAC</u>	<u>1</u>	<u>N/C W/ARR</u>
<u>MPLABOR</u>		<u>.5</u>	<u>N/C W/ARR</u>

Repair Notes:

REPLACE FC 1

CAL W/QC & OPS CHECKS

Service Technician: [Signature]

Date: 02/04/2009