

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Carney's Point P.D., Carneys Point, NJ

Date Received: 04/12/12 Date given to service: 04/12/12

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810 7110 7510 7410 Upper-half 7410 Whole
 8610 DT5000 9510

Serial Number: AR TN-0055
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>Fuel Cell</u>	<u>1</u>	<u>n/c w</u>
<u>mp Cal 71</u>	<u>Calibration</u>	<u>1</u>	<u>n/c w</u>
<u>mp Labor</u>	<u>Labor</u>	<u>.5</u>	<u>n/c w</u>

Repair Notes: Replaced Fuel cell # AR3K-0662, because of Leakage. New Fuel Cell # ARCM-1583.

Cal w/DC ; ops check

Service Technician: [Signature] Date: 4-13-12