

# Return & Repair Form:

## Customer Information:

Company Name: DELRAN TWP B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 04/28/2009 Date Given to Service: 04/28/2009  
 Carrier: PedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410 - 7110 - 6510 - 6810 Serial #: ARTN - 0057  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -

## ACCESSORIES:

110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

## Test #

Reason for Return: \_\_\_\_\_

Part Number	Description	Qty	Total Cost
<u>16808455</u>	<u>FC</u>	<u>1</u>	<u>N/C w/inv</u>
<u>MP cal 71</u>	<u>Cal</u>	<u>1</u>	<u>N/C w/inv</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C w/inv</u>

## Repair Notes:

Replace FC  
cal w/QC & ops check

Service Technician: [Signature]

Date: 04/29/2009