

Dräger safety

Internal Use Only

Return & Repair Form

Customer Information: LINCOLN PARK PD

CUSTOMER BARBARA WILSON XT7154 Cust # 417261

Date of Return: 12/22 Carrier: FedEx Method: 2 DAY

Product Information:

Product: 7110 7410 6510 6810 Serial # ACTN-0064

Description: A B PLUS SCREENER _ Printer Serial # _____

Whole Inst. Top Half Simulator Serial # _____

OTHER: _____ Probe Serial # _____

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter
- Regulator
- Mag Card Rdr # _____
- Printer Paper
- Printer Ribbon
- Casio # _____
- Mouthpieces
- Carrying Case
- Dry Gas
- Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: Power shuts off after powered on.

Part #	Description	Qty	Total
<u>8315075</u>	<u>Mother board</u>	<u>1</u>	
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Replaced the mother board. Cal, Ops Check & O.C.

Technician: CRD

Date: 12/27/05