

# Dräger safety

## Return & Repair Form

### Customer Information:

CUSTOMER READINGTON TWP. Police Dept. Cust # \_\_\_\_\_  
 Date of Return: 6-14-7 Carrier: Fed Ex Method: Exp. Saver

### Product Information:

**Product:** 7110 7410 6510 6810 Serial # ARTN-0067  
**Description:** A B PLUS SCREENER Printer Serial # \_\_\_\_\_  
**Whole Inst. Top Half** Simulator Serial # \_\_\_\_\_  
**OTHER:** \_\_\_\_\_ Probe Serial # \_\_\_\_\_  
 Demo Unit  Returned to stock on: \_\_\_\_\_  
**Accessories:** (check all that apply):  
 110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (specify): \_\_\_\_\_  
 Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: EC Problems

Part #	Description	Qty	Total
<u>6808455</u>	<u>Fuel Cell</u>	<u>1</u>	
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	
<u>MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	

Repair Notes: Replaced leaky F/C. Full cal ops check + O.C.

Technician: CRD

Date: 6/18/07