

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Readington Twp. P.D., Whitehouse Station, NJ

Date Received: 03/19/15

Date given to service: 03/19/15

Carrier: FedEx UPS USPS

Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR TN-0067

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: October 2016

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	2	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W
MP Misc.	Power Supply	1	NC/W

Repair Notes: Replaced defective power supply, fuel cell, and spacer plates.

Unit was not powering up properly by changing the power supply resolved the issue.

Fuel Cell = ARFJ-0641

CAL W/QC AND OPS CHECK

Service Technician: BC

Date: 03/25/15