

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER Washington Twp Police Cust # _____
 Date of Return: 7/15 Carrier: Fed Exp Method: 20AY

Product Information:

Product: 7110 7410 6510 Printer Simulator Probe Serial # ARTN-0069
Description: A B PLUS SCREENER Serial # _____
Whole Inst. Top Half Printer Serial # _____
OTHER: _____ Probe Serial # DDUJPZ-060
 Demo Unit Returned to stock on: _____
Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: will NOT PRINT A COPY OF TEST

Part #	Description	Qty	Total
<u>8314351</u>	<u>PCB (WARRANTY)</u>	<u>1</u>	
<u>MPCAL71</u>	<u>CAL WARRANTY</u>	<u>1</u>	

Repair Notes: "Copy" will not work due to data corruption - REPLACE PCB

PROBE FAILED QC - RETURN TO STOCK & REPLACE

Technician: [Signature]

Date: 07.17.05