

# Drägersafety

Internal Use Only

## Return & Repair Form

### Customer Information:

CUSTOMER Lumberton Twp. Police Cust # \_\_\_\_\_  
Date of Return: 9/28/06 Carrier: FED EX Method: Priority

### Product Information:

Product: 7110 7410 6510 6810 Serial # ARUL-0042

Description: A B PLUS SCREENER \_ Printer Serial # \_\_\_\_\_

Whole Inst. Top Half Simulator Serial # \_\_\_\_\_

OTHER: \_\_\_\_\_ Probe Serial # \_\_\_\_\_

Demo Unit  Returned to stock on: \_\_\_\_\_

### Accessories: (check all that apply):

110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_

Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_

Mouthpieces  Carrying Case  Dry Gas

Other (specify): \_\_\_\_\_

Warranty Expiration Date: \_\_\_\_\_

### Repair Information: Test # \_\_\_\_\_

Reason for Return: Ambient Air Check Error

W  
W  
W

Part #	Description	Qty	Total
<u>6808455</u>	<u>Fuel Cell</u>	<u>1</u>	
<u>MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Replaced F/C. It dropped low upon 75 tests. Full cal ops check + Q.C.

Technician: CRD

Date: 10/4/06