

RETURN AND REPAIR FORM



Customer Information

Company Name: N.J.S.P., Mt. Laurel, N.J.

Date Received: 04/10/12 Date given to service: 04/10/12

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR ULH-0044
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

| Part Number | Description | Qty | Total Cost |
|------------------|--------------------|-----------|--------------|
| <u>6808455</u> | <u>Fuel Cell</u> | <u>1</u> | <u>N/C W</u> |
| <u>mp Cal 71</u> | <u>Calibration</u> | <u>1</u> | <u>N/C W</u> |
| <u>mp Labor</u> | <u>Labor</u> | <u>15</u> | <u>N/C W</u> |

Repair Notes: Replaced Fuel Cell # ARZK-0828 with Fuel Cell # ARcm-0888.

Cal w/ Oc & ops Check

Service Technician: [Signature] Date: 4-12-12