

RETURN AND REPAIR FORM



Customer Information

Company Name: New Jersey State Police, Mt. Laurel, NJ

Date Received: 09/17/2016

Date given to service: 09/17/2016

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT6000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR UL-0044

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade-In

Accessories

110V A/C Adaptor Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information

Part Number	Description	Quantity	NC/W
4414161	Calibration	1	NC/W
4414166	Labor	.5	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W

Repair Notes: Replaced defective fuel cell and spacer plates.

Fuel Cell = ARJE-1931

CAL W/QC AND OPS CHECK

Service Technician: BC

Date: 10/14/2016