

# Return & Repair Form:

## Customer Information:

PD

Company Name: Rockaway Township S: \_\_\_\_\_

Date Received: \_\_\_\_\_ Date Given to Service: \_\_\_\_\_

Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

*EXPRESS Saver*

## Product Information:

Product: 7410 - 7110 - 6510 - 6810 Serial #: AR UL - 0045

Description: A - B - Plus - Screener - Demo Printer Ser #: AR - \_\_\_\_\_

Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_

Other \_\_\_\_\_ Probe Ser # DD P - \_\_\_\_\_

## ACCESSORIES:

110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_

Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_

Mouthpieces  Carrying Case  Dry Gas

Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return: FC

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP Cal 71</u>	<u>Cal</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP Labor</u>	<u>Labor</u>	<u>1.5</u>	<u>N/C WARR</u>

Repair Notes: REPLACE FC

Cal w/Op & Ops Check

Service Technician: [Signature]

Date: 7-8-2009